



# Immunosuppressant (IPT) Individual Report

Round: IP426

IP0402 - Allgemeines Krankenhaus Wien

Issue Number: 1

Issued: 11/10/2019



**Science**  
for a safer world

**LGC Standards Proficiency Testing**  
1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

**CICTAC - Ciclosporin and Tacrolimus**

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Ciclosporin	Lab Result	HPLC-MS	115.00	µg/L	-0.25	117.00	0.91	8.15	N/A	127	117.00	117.84	8.149	9.321	HPLC-MS
B - Ciclosporin	Lab Result	HPLC-MS	118.00	µg/L	0.00	118.00	0.99	8.9	N/A	127	118.00	118.44	8.898	9.797	HPLC-MS
C - Ciclosporin	Lab Result	HPLC-MS	124.00	µg/L	0.81	118.00	0.82	7.42	N/A	127	118.00	118.93	7.415	9.876	HPLC-MS
A - Tacrolimus	Lab Result	HPLC-MS	8.20	µg/L	0.17	8.10	0.06	0.59	N/A	139	8.10	8.13	0.593	0.599	HPLC-MS
B - Tacrolimus	Lab Result	HPLC-MS	2.30	µg/L	1.16	2.11	0.02	0.16	N/A	139	2.11	2.15	0.163	0.210	HPLC-MS
C - Tacrolimus	Lab Result	HPLC-MS	20.50	µg/L	0.39	20.10	0.11	1.04	N/A	139	20.10	20.20	1.038	1.224	HPLC-MS

**EVE - Everolimus**

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Everolimus	Lab Result	HPLC-MS	8.00	µg/L	0.00	8.00	0.09	0.74	N/A	105	8.00	8.10	0.742	0.898	HPLC-MS
B - Everolimus	Lab Result	HPLC-MS	8.20	µg/L	0.25	8.00	0.10	0.79	N/A	105	8.00	8.06	0.793	0.919	HPLC-MS
C - Everolimus	Lab Result	HPLC-MS	7.80	µg/L	-0.25	8.00	0.10	0.8	N/A	105	8.00	8.17	0.801	0.998	HPLC-MS

**SIR- Sirolimus**

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Sirolimus	Lab Result	HPLC-MS	10.60	µg/L	-0.04	10.64	0.13	1.1	N/A	120	10.64	10.67	1.097	1.230	HPLC-MS
B - Sirolimus	Lab Result	HPLC-MS	7.70	µg/L	0.40	7.40	0.09	0.74	N/A	120	7.40	7.40	0.742	0.881	HPLC-MS
C - Sirolimus	Lab Result	HPLC-MS	4.10	µg/L	-0.59	4.46	0.07	0.6	N/A	120	4.46	4.47	0.601	0.663	HPLC-MS