



# Immunosuppressant (IPT) Individual Report

Round: IP429 - (Round 429)

IP0402 - Allgemeines Krankenhaus Wien

Issue Number: 1

Issued: 09/01/2020



Science  
for a safer world

**LGC Standards Proficiency Testing**  
1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

**CICTAC - Ciclosporin and Tacrolimus**

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Ciclosporin	Lab Result	HPLC-MS	1033.00	µg/L	0.74	977.50	8.56	75.0	N/A	122	977.50	976.38	75.003	82.670	HPLC-MS
B - Ciclosporin	Lab Result	HPLC-MS	395.00	µg/L	1.15	366.00	2.86	25.21	N/A	122	366.00	363.20	25.211	29.654	HPLC-MS
C - Ciclosporin	Lab Result	HPLC-MS	1256.00	µg/L	-0.11	1267.50	12.48	107.52	N/A	122	1267.50	1269.26	107.518	107.013	HPLC-MS
A - Tacrolimus	Lab Result	HPLC-MS	7.10	µg/L	-0.67	7.42	0.05	0.47	N/A	134	7.42	7.47	0.475	0.577	HPLC-MS
B - Tacrolimus	Lab Result	HPLC-MS	2.80	µg/L	-0.52	2.90	0.02	0.19	N/A	134	2.90	2.89	0.193	0.269	HPLC-MS
C - Tacrolimus	Lab Result	HPLC-MS	13.70	µg/L	-0.67	14.20	0.08	0.74	N/A	134	14.20	14.14	0.742	0.842	HPLC-MS

**SIR- Sirolimus**

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Sirolimus	Lab Result	HPLC-MS	4.60	µg/L	0.89	4.19	0.05	0.46	N/A	114	4.19	4.19	0.460	0.505	HPLC-MS
B - Sirolimus	Lab Result	HPLC-MS	4.40	µg/L	0.44	4.20	0.05	0.46	N/A	114	4.20	4.21	0.460	0.524	HPLC-MS
C - Sirolimus	Lab Result	HPLC-MS	4.80	µg/L	1.25	4.17	0.06	0.5	N/A	114	4.17	4.18	0.497	0.557	HPLC-MS