



Immunosuppressant (IPT) Individual Report

IP438 - (Round 438) 05 Oct 2020

IP0402 - Universitätsklinikum AKH Wien

Issue Number: 1

Issued: 14/10/2020



Science
for a safer world

LGC Proficiency Testing
1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

CICTAC - Ciclosporin and Tacrolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Ciclosporin	Lab Result	HPLC-MS	126.00	µg/L	0.67	120.00	0.99	8.9	N/A	129	120.00	120.66	8.898	13.014	HPLC-MS
B - Ciclosporin	Lab Result	HPLC-MS	122.00	µg/L	0.15	120.60	1.06	9.64	N/A	129	120.60	121.48	9.640	11.443	HPLC-MS
C - Ciclosporin	Lab Result	HPLC-MS	118.00	µg/L	-0.29	121.00	1.14	10.31	N/A	129	121.00	120.68	10.307	11.444	HPLC-MS
A - Tacrolimus	Lab Result	HPLC-MS	5.00	µg/L	1.35	4.60	0.03	0.3	N/A	141	4.60	4.57	0.297	0.393	HPLC-MS
B - Tacrolimus	Lab Result	HPLC-MS	4.40	µg/L	-0.67	4.60	0.03	0.3	N/A	141	4.60	4.59	0.297	0.368	HPLC-MS
C - Tacrolimus	Lab Result	HPLC-MS	4.50	µg/L	0.00	4.50	0.03	0.3	N/A	141	4.50	4.57	0.297	0.389	HPLC-MS

EVE - Everolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Everolimus	Lab Result	HPLC-MS	9.40	µg/L	-0.27	9.66	0.12	0.98	N/A	98	9.66	9.56	0.979	1.164	HPLC-MS
B - Everolimus	Lab Result	HPLC-MS	2.30	µg/L	1.19	2.00	0.03	0.25	N/A	98	2.00	2.02	0.252	0.322	HPLC-MS
C - Everolimus	Lab Result	HPLC-MS	7.80	µg/L	0.38	7.51	0.10	0.76	N/A	98	7.51	7.57	0.756	1.097	HPLC-MS

SIR- Sirolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Sirolimus	Lab Result	HPLC-MS	9.00	µg/L	0.23	8.79	0.11	0.9	N/A	115	8.79	8.83	0.905	1.042	HPLC-MS
B - Sirolimus	Lab Result	HPLC-MS	7.00	µg/L	0.00	7.00	0.09	0.74	N/A	115	7.00	7.04	0.742	0.846	HPLC-MS
C - Sirolimus	Lab Result	HPLC-MS	7.10	µg/L	0.15	7.00	0.08	0.65	N/A	115	7.00	7.09	0.653	0.890	HPLC-MS