



Immunosuppressant (IPT) Individual Report

IP441 - (Round 441) 11 Jan 2021

IP0402 - Universitätsklinikum AKH Wien

Issue Number: 1

Issued: 13/01/2021



Science
for a safer world

LGC Proficiency Testing
1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

CICTAC - Ciclosporin and Tacrolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Ciclosporin	Lab Result	HPLC-MS	40.00	µg/L	1.93	33.64	0.37	3.29	N/A	125	33.64	33.65	3.292	3.463	HPLC-MS
B - Ciclosporin	Lab Result	HPLC-MS	158.00	µg/L	1.34	146.04	1.00	8.95	N/A	124	146.04	146.83	8.950	12.338	HPLC-MS
C - Ciclosporin	Lab Result	HPLC-MS	1453.00	µg/L	0.82	1360.61	13.11	113.0	N/A	124	1360.61	1357.54	112.997	124.024	HPLC-MS
A - Tacrolimus	Lab Result	HPLC-MS	7.50	µg/L	-0.22	7.60	0.05	0.44	N/A	132	7.60	7.66	0.445	0.612	HPLC-MS
B - Tacrolimus	Lab Result	HPLC-MS	3.90	µg/L	-0.54	4.06	0.03	0.3	N/A	132	4.06	4.08	0.304	0.385	HPLC-MS
C - Tacrolimus	Lab Result	HPLC-MS	18.70	µg/L	0.51	18.10	0.13	1.19	N/A	132	18.10	18.22	1.186	1.498	HPLC-MS

SIR- Sirolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Sirolimus	Lab Result	HPLC-MS	11.20	µg/L	0.99	10.19	0.12	1.02	N/A	110	10.19	10.30	1.023	1.307	HPLC-MS
B - Sirolimus	Lab Result	HPLC-MS	10.30	µg/L	0.00	10.30	0.12	1.04	N/A	110	10.30	10.36	1.038	1.257	HPLC-MS
C - Sirolimus	Lab Result	HPLC-MS	11.00	µg/L	0.97	10.10	0.11	0.93	N/A	110	10.10	10.25	0.927	1.317	HPLC-MS