



Immunosuppressant (IPT) Individual Report

IP453 - (Round 453) 24 Jan 2022

IP0402 - Universitätsklinikum AKH Wien

Issue Number: 1

Issued: 26/01/2022



Science
for a safer world

LGC Proficiency Testing
1 Chamberhall Business Park | Chamberhall Green | Bury | United Kingdom | BL9 0AP



0001

CICTAC - Ciclosporin and Tacrolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Ciclosporin	Lab Result	HPLC-MS	1499.00	µg/L	0.59	1423.70	14.562	127.612	N/A	127	1423.70	1408.40	127.612	140.996	HPLC-MS
B - Ciclosporin	Lab Result	HPLC-MS	338.00	µg/L	1.43	304	2.632	23.728	N/A	127	304	305.92	23.728	24.417	HPLC-MS
C - Ciclosporin	Lab Result	HPLC-MS	139.00	µg/L	0.87	130	1.156	10.381	N/A	126	130	129.83	10.381	11.123	HPLC-MS
A - Tacrolimus	Lab Result	HPLC-MS	27.80	µg/L	-0.27	28.20	0.161	1.483	N/A	133	28.20	28.15	1.483	1.646	HPLC-MS
B - Tacrolimus	Lab Result	HPLC-MS	7.80	µg/L	0.17	7.72	0.051	0.475	N/A	133	7.72	7.71	0.475	0.553	HPLC-MS
C - Tacrolimus	Lab Result	HPLC-MS	20.60	µg/L	0.96	19.66	0.106	0.979	N/A	133	19.66	19.62	0.979	1.131	HPLC-MS

SIR- Sirolimus

Analyte	Analyst	Method	Result	Units	Z Score	Assigned Value	Ux AV	SDPA	Exp.SDPA	Number of results	Median	Mean	Robust SD	SD	Your Reference
A - Sirolimus	Lab Result	HPLC-MS	21.60	µg/L	-0.87	23.15	0.212	1.78	N/A	112	23.15	22.93	1.780	2.207	HPLC-MS
B - Sirolimus	Lab Result	HPLC-MS	12.30	µg/L	0.00	12.30	0.124	1.038	N/A	112	12.30	12.31	1.038	1.217	HPLC-MS
C - Sirolimus	Lab Result	HPLC-MS	5.30	µg/L	0.63	5.06	0.044	0.371	N/A	112	5.06	5.03	0.371	0.516	HPLC-MS