

EINGEGANGEN
20. Aug. 2013
TQM

[Handwritten signature]

International proficiency testing scheme - Everolimus

Centre 4021



PT Distribution 73

Samples dispatched on: 9th July 2013
Results deadline: 16th August 2013

Method: HPLC-MS or HPLC-MS/MS

Group Code: HPLC

Bias plot: Distribution 70 - 73

<p>PT Code 73A</p>	<p>Result</p> <p>Target concentration (ng/mL) 1.0 Method mean (ng/mL) 1.1 SD 0.3 CV 27.3% Your result (ng/mL) <3 ✓ Inaccuracy Z Value Accepted Yes Total number of results 111 Number of accepted results 107</p>	
<p>PT Code 73B</p>	<p>Result</p> <p>Target concentration (ng/mL) 15.0 Method mean (ng/mL) 15.2 SD 1.5 CV 9.9% Your result (ng/mL) 14.3 ✓ Inaccuracy Z Value -0.6 Accepted Yes Total number of results 111 Number of accepted results 111</p>	
<p>PT Code 73C</p>	<p>Result</p> <p>Target concentration (ng/mL) 3.0 Method mean (ng/mL) 3.0 SD 0.4 CV 13.3% Your result (ng/mL) <3 ✓ Inaccuracy Z Value Accepted Yes Total number of results 111 Number of accepted results 107</p>	

PT Code	70A	70B	70C	71A	71B	71C	72A	72B	72C	73A	73B	73C
Bias	5.1%	15.4%	9.1%	3.3%	6.5%	6.3%	8.1%	8.1%	1.6%		-5.9%	
Target value	15.0	3.0	5.0	Pool	10.0	4.0	6.0	6.0	6.0	1.0	15.0	3.0
Method Mean	13.6	2.6	4.4	6.0	9.3	3.2	6.2	6.2	6.2	1.1	15.2	3.0
Your value	14.3	3.0	4.8	6.2	9.9	3.4	6.7	6.7	6.3	<3	14.3	<3
Accepted	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes

[Handwritten signature]