

EINGEGANGEN
02. Juli 2015
TQM

Centre 145

International proficiency testing scheme - Ciclosporin







PT Distribution 375

Samples dispatched on: 02 June 2015
Results deadline: 26th June 2015

Method: **ACMIA** – on Dimension platform

Group Code: **ACMIA** – on Dimension platform

Bias plot: **Distribution 372 - 375**



PT Code	372A	372B	372C	373A	373B	373C	374A	374B	374C	375A	375B	375C
Bias	-2.7%	0.3%	4.3%	-2.0%	-3.1%	-3.7%	0.0%	-7.0%	8.1%	-3.2%	-1.2%	
Target value	100.0	1250.0	300.0	750.0	Pool	400.0	80.0	1000.0	Pool	Pool	1500.0	Blank
Method Mean	96.6	1289.6	299.2	781.6	167.2	403.0	79.0	1077.0	142.4	155.0	1557.7	4.6
Your value	94.0	1293.0	312.0	765.0	162.0	388.0	79.0	1002.0	154.0	150.0	1539.0	<30
Accepted	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes