



# **Immunosuppressant (IPT) IP0402 - Allgemeines Krankenhaus Wien Individual Report**

**Round: 394**

Issue Number 1

Issued 10 February 2017



**Science**  
for a safer world

## CICTAC - Ciclosporin and/or Tacrolimus

Analyte	Analyst	Method	Result	Units	z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
A - Ciclosporin	Lab Result	HPLC-MS	<26.10	µg/L		Zero Spike	N/A	N/A	N/A	167	N/A	N/A	N/A	N/A
B - Ciclosporin	Lab Result	HPLC-MS	658.00	µg/L	0.65	633.00	3.75	38.558	N/A	168	633.00	627.58	38.558	48.060
C - Ciclosporin	Lab Result	HPLC-MS	37.00	µg/L	1.10	33.90	0.27	2.818	N/A	168	33.90	33.49	2.818	4.116
A - Tacrolimus	Lab Result	HPLC-MS	15.00	µg/L	0.76	14.10	0.11	1.186	N/A	176	14.10	14.24	1.186	1.329
B - Tacrolimus	Lab Result	HPLC-MS	<2.40	µg/L		Zero Spike	N/A	N/A	N/A	175	N/A	N/A	N/A	N/A
C - Tacrolimus	Lab Result	HPLC-MS	6.70	µg/L	0.51	6.40	0.06	0.593	N/A	177	6.40	6.46	0.593	0.613

## SIR- Sirolimus

Analyte	Analyst	Method	Result	Units	z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
A - Sirolimus	Lab Result	HPLC-MS	9.70	µg/L	0.24	9.45	0.11	1.046	N/A	140	9.45	9.45	1.046	1.190
B - Sirolimus	Lab Result	HPLC-MS	4.90	µg/L	0.71	4.44	0.07	0.645	N/A	139	4.44	4.51	0.645	0.648
C - Sirolimus	Lab Result	HPLC-MS	23.50	µg/L	1.19	20.50	0.27	2.521	N/A	139	20.50	20.95	2.521	2.828

## EVE - Everolimus

Analyte	Analyst	Method	Result	Units	z score (** z' score)	Assigned Value	Ux AV	SDPA	Exp.SDPA	No of results	Median	Mean	Robust SD	SD
A - Everolimus	Lab Result	HPLC-MS	<2.30	µg/L		Zero Spike	N/A	N/A	N/A	129	N/A	N/A	N/A	N/A
B - Everolimus	Lab Result	HPLC-MS	<2.30	µg/L		2.10	0.05	0.437	N/A	129	2.10	2.12	0.437	0.400
C - Everolimus	Lab Result	HPLC-MS	13.10	µg/L	1.35	11.70	0.11	1.038	N/A	129	11.70	11.80	1.038	1.346

\*\* Please note, participant performance for this analyte has been assessed using a z' score, rather than a z score, in order to account for the measurement uncertainty of the assigned value which is not negligible when compared to the SDPA.